

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004412.D
 Acq On : 07 Sep 2018 17:25
 Operator : JC/MD
 Sample : J4812-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Sep 08 04:14:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	322360	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	509387	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	483573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	268536	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	219446	52.23	ug/l	0.00
Spiked Amount						
			Recovery	=	104.46%	
35) Dibromofluoromethane	5.50	113	194344	51.28	ug/l	0.00
Spiked Amount						
			Recovery	=	102.56%	
50) Toluene-d8	8.72	98	721988	50.47	ug/l	0.00
Spiked Amount						
			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	11.14	95	258452	48.42	ug/l	0.00
Spiked Amount						
			Recovery	=	96.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	12187	2.89	ug/l	# 83
16) Acetone	2.45	43	14357	0.31	ug/l	99
19) Methyl tert-butyl Ether	3.20	73	70425	6.26	ug/l	96
21) trans-1,2-Dichloroethene	3.16	96	1326	0.35	ug/l	93
27) cis-1,2-Dichloroethene	4.60	96	86360	20.14	ug/l	89
44) Trichloroethene	7.21	130	45536	9.61	ug/l	98
64) Tetrachloroethene	9.34	164	4822	1.01	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004412.D
Acq On : 07 Sep 2018 17:25
Operator : JC/MD
Sample : J4812-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Sep 08 04:14:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

