

Data File : VX004415.D
Acq On : 07 Sep 2018 18:45
Operator : JC/MD
Sample : J4812-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-8

Quant Time: Sep 14 07:05:35 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	334871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	523562	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	509045	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	288154	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	221059	50.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.30%	
35) Dibromofluoromethane	5.50	113	195935	50.30	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.60%	
50) Toluene-d8	8.72	98	770304	52.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.78%	
62) 4-Bromofluorobenzene	11.14	95	283402	51.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.32%	
Target Compounds						
31) Cyclohexane	5.57	56	73073	11.941	ug/l	95
39) Methylcyclohexane	7.46	83	45500	6.771	ug/l	89
73) Isopropylbenzene	11.02	105	19436	1.027	ug/l	100
78) n-propylbenzene	11.36	91	17600	0.809	ug/l	99
85) sec-Butylbenzene	11.94	105	6232	0.333	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX004415.D

Acq On : 07 Sep 2018 18:45

Operator : JC/MD

Sample : J4812-07

Misc : 5.0mL/MSVOA_X/WATER

ALS Vial : 16 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

MW-8

Quant Time: Sep 14 07:05:35 2018

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M

Quant Title : SW846 8260

QLast Update : Thu Aug 23 05:39:44 2018

Response via : Initial Calibration

