

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004431.D
 Acq On : 08 Sep 2018 02:45
 Operator : JC/MD
 Sample : PB112672ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB112672ZHE#07

Quant Time: Sep 08 05:50:18 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	293865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	490699	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	451208	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	261782	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	210473	54.95	ug/l	0.00
Spiked Amount			Recovery	=	109.90%	
35) Dibromofluoromethane	5.50	113	184922	50.66	ug/l	0.00
Spiked Amount			Recovery	=	101.32%	
50) Toluene-d8	8.72	98	693005	50.29	ug/l	0.00
Spiked Amount			Recovery	=	100.58%	
62) 4-Bromofluorobenzene	11.14	95	244554	47.57	ug/l	0.00
Spiked Amount			Recovery	=	95.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	64581	34.34	ug/l	96
18) Methyl Acetate	2.78	43	13528	2.38	ug/l	97
20) Methylene Chloride	2.85	84	65600	17.79	ug/l	95
43) Isopropyl Acetate	6.46	43	207948	24.94	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004431.D
Acq On : 08 Sep 2018 02:45
Operator : JC/MD
Sample : PB112672ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB112672ZHE#07

Quant Time: Sep 08 05:50:18 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

