

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004433.D
 Acq On : 08 Sep 2018 03:39
 Operator : JC/MD
 Sample : J4779-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-03-090618-B

Quant Time: Sep 08 05:57:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	299662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	502085	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	475381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	276259	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	214105	54.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.64%	
35) Dibromofluoromethane	5.50	113	188934	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
50) Toluene-d8	8.71	98	728785	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.14	95	259278	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	23311	6.77	ug/l	97
18) Methyl Acetate	2.78	43	13658	2.36	ug/l	97
20) Methylene Chloride	2.85	84	1932004	513.87	ug/l	98
43) Isopropyl Acetate	6.46	43	169381	19.86	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004433.D
Acq On : 08 Sep 2018 03:39
Operator : JC/MD
Sample : J4779-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-03-090618-B

Quant Time: Sep 08 05:57:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

