

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090718\
 Data File : VX004436.D
 Acq On : 08 Sep 2018 04:59
 Operator : JC/MD
 Sample : J4774-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OK-03-090618

Quant Time: Sep 08 06:06:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	295335	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	488407	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	449132	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	264062	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	210757	54.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.50%	
35) Dibromofluoromethane	5.50	113	187517	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
50) Toluene-d8	8.71	98	703624	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.14	95	248585	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
Target Compounds						
16) Acetone	2.44	43	25357	8.33	ug/l	97
18) Methyl Acetate	2.78	43	11296	1.98	ug/l	97
20) Methylene Chloride	2.85	84	122787	33.14	ug/l	97
30) Chloroform	5.21	83	11540	1.84	ug/l	91
43) Isopropyl Acetate	6.46	43	213757	25.76	ug/l	99
44) Trichloroethene	7.21	130	4892	1.08	ug/l	99
95) Naphthalene	13.83	128	227109	13.18	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090718\
Data File : VX004436.D
Acq On : 08 Sep 2018 04:59
Operator : JC/MD
Sample : J4774-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
OK-03-090618

Quant Time: Sep 08 06:06:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

