

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090719\
 Data File : VX012170.D
 Acq On : 07 Sep 2019 10:30
 Operator : SY/SP
 Sample : VX0907SBL01
 Misc : 5.00g/5.0mL/MSVOA X/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0907SBL01

Quant Time: Sep 08 07:59:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	152655	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	265158	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	234327	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	112659	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	102063	48.46	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.92%	
35) Dibromofluoromethane	5.48	113	82766	50.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.52%	
50) Toluene-d8	8.71	98	319452	53.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.78%	
62) 4-Bromofluorobenzene	11.13	95	132823	47.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090719\
Data File : VX012170.D
Acq On : 07 Sep 2019 10:30
Operator : SY/SP
Sample : VX0907SBL01
Misc : 5.00µ/5.0mL/MSVOA X/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0907SBL01

Quant Time: Sep 08 07:59:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

