

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090719\
 Data File : VX012182.D
 Acq On : 07 Sep 2019 17:59
 Operator : SY/SP
 Sample : K4602-11
 Misc : 4.75g/5.0mL/MSVOA X/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-C3-(3.11)

Quant Time: Sep 08 08:24:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	152760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	267405	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	233879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	99583	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	108892	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
35) Dibromofluoromethane	5.48	113	80109	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
50) Toluene-d8	8.71	98	317908	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.14	95	125659	44.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.56%	
Target Compounds						
20) Methylene Chloride	2.84	84	6531	3.228	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090719\
Data File : VX012182.D
Acq On : 07 Sep 2019 17:59
Operator : SY/SP
Sample : K4602-11
Misc : 4.75g/5.0mL/MSVOA X/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-C3-(3.11)

Quant Time: Sep 08 08:24:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

