

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090719\
 Data File : VX012183.D
 Acq On : 07 Sep 2019 18:27
 Operator : SY/SP
 Sample : K4602-12
 Misc : 5.68g/5.0mL/MSVOA X/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-C4-(3.11)

Quant Time: Sep 08 08:25:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	145750	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	251944	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	226287	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	103221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	101899	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
35) Dibromofluoromethane	5.48	113	76207	49.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.38%	
50) Toluene-d8	8.71	98	303638	53.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.82%	
62) 4-Bromofluorobenzene	11.13	95	124406	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
Target Compounds						
20) Methylene Chloride	2.84	84	8080	4.186	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090719\
Data File : VX012183.D
Acq On : 07 Sep 2019 18:27
Operator : SY/SP
Sample : K4602-12
Misc : 5.68g/5.0mL/MSVOA X/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-C4-(3.11)

Quant Time: Sep 08 08:25:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

