

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090719\
 Data File : VX012186.D
 Acq On : 07 Sep 2019 19:51
 Operator : SY/SP
 Sample : K4601-09
 Misc : 4.82g/5.0mL/MSVOA X/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T12-E4-(2.7)

Quant Time: Sep 08 08:29:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	143244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	250634	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	209162	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	78874	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	100722	50.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.92%	
35) Dibromofluoromethane	5.48	113	74283	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	290255	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
62) 4-Bromofluorobenzene	11.13	95	106307	40.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.84%	
Target Compounds						
16) Acetone	2.43	43	1711	3.622	ug/l	95
20) Methylene Chloride	2.84	84	5576	2.939	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090719\
Data File : VX012186.D
Acq On : 07 Sep 2019 19:51
Operator : SY/SP
Sample : K4601-09
Misc : 4.82q/5.0mL/MSVOA X/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T12-E4-(2.7)

Quant Time: Sep 08 08:29:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

