

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031157.D
 Acq On : 07 Sep 2022 13:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Sep 07 23:24:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	85829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	148015	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127967	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56814	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	62040	30.531	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	61.060%#
35) Dibromofluoromethane	5.391	113	47626	29.527	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	59.060%#
50) Toluene-d8	8.653	98	172516	29.438	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	58.880%#
62) 4-Bromofluorobenzene	11.079	95	65647	30.568	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	61.140%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031157.D
Acq On : 07 Sep 2022 13:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Sep 07 23:24:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

