

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031161.D
 Acq On : 07 Sep 2022 15:56
 Operator : JC/MD
 Sample : VX0907WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0907WBL01

Quant Time: Sep 07 23:25:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	145193	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	114100	57.655	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.320%
35) Dibromofluoromethane	5.385	113	87554	55.336	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.680%
50) Toluene-d8	8.647	98	309022	53.757	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	107.520%
62) 4-Bromofluorobenzene	11.079	95	116998	55.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	111.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031161.D
Acq On : 07 Sep 2022 15:56
Operator : JC/MD
Sample : VX0907WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0907WBL01

Quant Time: Sep 07 23:25:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

