

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031166.D
 Acq On : 07 Sep 2022 18:02
 Operator : JC/MD
 Sample : N4357-16DL 2X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP05-20220823DL

Quant Time: Sep 07 23:26:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	84920	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	152663	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	137616	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55631	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113672	56.538	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 113.080%		
35) Dibromofluoromethane	5.391	113	86877	52.222	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.440%		
50) Toluene-d8	8.653	98	310272	51.333	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.660%		
62) 4-Bromofluorobenzene	11.079	95	116284	52.498	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 105.000%		
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	33177	21.749	ug/l	96
27) cis-1,2-Dichloroethene	4.495	96	796	0.436	ug/l	88
44) Trichloroethene	7.129	130	65864	34.621	ug/l	89
64) Tetrachloroethene	9.281	164	3503	2.445	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031166.D
Acq On : 07 Sep 2022 18:02
Operator : JC/MD
Sample : N4357-16DL 2X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP05-20220823DL

Quant Time: Sep 07 23:26:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

