

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031167.D
 Acq On : 07 Sep 2022 18:26
 Operator : JC/MD
 Sample : N4357-14DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20220823DL

Quant Time: Sep 07 23:26:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 09/08/2022
 Supervised By : Mahesh Dadoda 09/08/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	144469	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129494	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54465	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110536	56.400	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.800%			
35) Dibromofluoromethane	5.385	113	87249	55.420	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 110.840%			
50) Toluene-d8	8.653	98	299498	52.361	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.720%			
62) 4-Bromofluorobenzene	11.079	95	110034	52.493	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 104.980%			
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	762	0.512	ug/l #	92
27) cis-1,2-Dichloroethene	4.483	96	876m	0.493	ug/l	
44) Trichloroethene	7.129	130	38528	21.401	ug/l	91
64) Tetrachloroethene	9.275	164	1676	1.243	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031167.D
Acq On : 07 Sep 2022 18:26
Operator : JC/MD
Sample : N4357-14DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20220823DL

Quant Time: Sep 07 23:26:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 09/08/2022
Supervised By : Mahesh Dadoda 09/08/2022

