

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031168.D
 Acq On : 07 Sep 2022 18:49
 Operator : JC/MD
 Sample : N4357-03DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D4-20220823DL

Quant Time: Sep 07 23:26:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	83983	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	147409	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	132607	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54619	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109209	54.925	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.840%	
35) Dibromofluoromethane	5.391	113	82627	51.437	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.880%	
50) Toluene-d8	8.653	98	292765	50.163	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.320%	
62) 4-Bromofluorobenzene	11.079	95	107960	50.477	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.960%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	43032	28.525	ug/l	95
12) 1,1-Dichloroethene	2.319	96	365	0.257	ug/l	# 41
44) Trichloroethene	7.129	130	5688	3.096	ug/l	92
64) Tetrachloroethene	9.281	164	2676	1.938	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 07 23:26:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

