

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031170.D
 Acq On : 07 Sep 2022 19:36
 Operator : JC/MD
 Sample : N4357-07DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D3-20220823DL

Quant Time: Sep 07 23:26:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	79726	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	144100	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49700	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108748	57.613	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.220%	
35) Dibromofluoromethane	5.385	113	83463	53.151	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.300%	
50) Toluene-d8	8.653	98	281848	49.402	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.800%	
62) 4-Bromofluorobenzene	11.079	95	104748	50.100	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.200%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	43610	30.451	ug/l	98
12) 1,1-Dichloroethene	2.319	96	471	0.350	ug/l	# 76
27) cis-1,2-Dichloroethene	4.489	96	423	0.247	ug/l	54
44) Trichloroethene	7.129	130	17185	9.570	ug/l	87
64) Tetrachloroethene	9.275	164	6772	5.242	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031170.D
Acq On : 07 Sep 2022 19:36
Operator : JC/MD
Sample : N4357-07DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D3-20220823DL

Quant Time: Sep 07 23:26:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

