

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031171.D
 Acq On : 07 Sep 2022 19:59
 Operator : JC/MD
 Sample : N4357-11DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW4-20220823DL

Quant Time: Sep 07 23:26:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	85688	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	149449	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135746	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56763	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110492	54.464	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.920%	
35) Dibromofluoromethane	5.391	113	84196	51.699	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.400%	
50) Toluene-d8	8.653	98	297912	50.348	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.700%	
62) 4-Bromofluorobenzene	11.079	95	111406	51.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.760%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.325	101	531	0.345	ug/l	# 86
44) Trichloroethene	7.129	130	69424	37.277	ug/l	95
64) Tetrachloroethene	9.281	164	467	0.330	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031171.D
Acq On : 07 Sep 2022 19:59
Operator : JC/MD
Sample : N4357-11DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW4-20220823DL

Quant Time: Sep 07 23:26:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

