

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031174.D
 Acq On : 07 Sep 2022 21:09
 Operator : JC/MD
 Sample : N4357-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04-20220822RE

Quant Time: Sep 07 23:27:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	82175	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	145927	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	131087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109246	56.152	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.300%
35) Dibromofluoromethane	5.385	113	83861	52.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.480%
50) Toluene-d8	8.647	98	279455	48.369	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.079	95	106091	50.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.220%
Target Compounds						
20) Methylene Chloride	2.794	84	682	0.431	ug/l	# 71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031174.D
Acq On : 07 Sep 2022 21:09
Operator : JC/MD
Sample : N4357-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB04-20220822RE

Quant Time: Sep 07 23:27:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

