

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031175.D
 Acq On : 07 Sep 2022 21:33
 Operator : JC/MD
 Sample : N4357-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE134D2-20220823RE

Quant Time: Sep 07 23:27:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81653	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	144484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52336	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	106751	55.221	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.440%	
35) Dibromofluoromethane	5.391	113	82274	52.255	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.500%	
50) Toluene-d8	8.653	98	281357	49.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.360%	
62) 4-Bromofluorobenzene	11.079	95	107854	51.448	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.900%	
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	13485	9.194	ug/l	100
12) 1,1-Dichloroethene	2.325	96	379	0.275	ug/l	# 45
27) cis-1,2-Dichloroethene	4.483	96	1746	0.996	ug/l	86
44) Trichloroethene	7.123	130	99014	54.992	ug/l	94
52) Toluene	8.720	92	2773	0.656	ug/l	86
64) Tetrachloroethene	9.275	164	9019	6.879	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031175.D
Acq On : 07 Sep 2022 21:33
Operator : JC/MD
Sample : N4357-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE134D2-20220823RE

Quant Time: Sep 07 23:27:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

