

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031177.D
 Acq On : 07 Sep 2022 22:19
 Operator : JC/MD
 Sample : N4357-17RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT150S1-20220823RE

Quant Time: Sep 07 23:27:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76816	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	141130	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125567	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113634	62.482	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.960%	
35) Dibromofluoromethane	5.391	113	82017	53.329	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.660%	
50) Toluene-d8	8.653	98	284914	50.990	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.980%	
62) 4-Bromofluorobenzene	11.079	95	108182	52.831	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.660%	
Target Compounds						
44) Trichloroethene	7.129	130	11685	6.644	ug/l	Qvalue 93
64) Tetrachloroethene	9.275	164	834	0.638	ug/l	# 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031177.D
Acq On : 07 Sep 2022 22:19
Operator : JC/MD
Sample : N4357-17RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT150S1-20220823RE

Quant Time: Sep 07 23:27:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

