

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031180.D
 Acq On : 07 Sep 2022 23:29
 Operator : JC/MD
 Sample : N4355-01DL 4X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-31DL

Quant Time: Sep 08 00:57:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	81887	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	144284	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130171	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110771	57.136	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.280%	
35) Dibromofluoromethane	5.391	113	84815	53.943	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.880%	
50) Toluene-d8	8.653	98	290429	50.841	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.680%	
62) 4-Bromofluorobenzene	11.079	95	114255	54.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.160%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.520	76	1333	0.392	ug/l	# 90
19) Methyl tert-butyl Ether	3.117	73	111527	22.712	ug/l	91
22) Diisopropyl ether	3.757	45	1582	0.406	ug/l	# 58
40) Benzene	6.043	78	19801	3.096	ug/l	98
52) Toluene	8.726	92	992	0.235	ug/l	88
67) Ethyl Benzene	10.195	91	10800	1.555	ug/l	# 85
68) m/p-Xylenes	10.305	106	6300	2.447	ug/l	91
69) o-Xylene	10.640	106	2615	1.018	ug/l	90
73) Isopropylbenzene	10.963	105	4216	0.855	ug/l	99
78) n-propylbenzene	11.305	91	5650	1.005	ug/l	97
80) 1,3,5-Trimethylbenzene	11.457	105	1816	0.438	ug/l	88
84) 1,2,4-Trimethylbenzene	11.756	105	6395	1.586	ug/l	89
95) Naphthalene	13.780	128	7571	1.691	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031180.D
Acq On : 07 Sep 2022 23:29
Operator : JC/MD
Sample : N4355-01DL 4X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-31DL

Quant Time: Sep 08 00:57:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

