

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031186.D
 Acq On : 08 Sep 2022 02:58
 Operator : JC/MD
 Sample : VX0907WBL02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0907WBL02

Quant Time: Sep 08 04:42:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78520	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	138929	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55021	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109949	59.144	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.280%
35) Dibromofluoromethane	5.391	113	82185	54.285	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.580%
50) Toluene-d8	8.653	98	287601	52.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.580%
62) 4-Bromofluorobenzene	11.079	95	109056	54.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031186.D
Acq On : 08 Sep 2022 02:58
Operator : JC/MD
Sample : VX0907WBL02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0907WBL02

Quant Time: Sep 08 04:42:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

