

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031192.D
 Acq On : 08 Sep 2022 05:41
 Operator : JC/MD
 Sample : N4446-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-04

Quant Time: Sep 08 13:58:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71339	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	136476	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121493	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53106	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99684	59.020	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.040%
35) Dibromofluoromethane	5.385	113	73069	49.131	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.260%
50) Toluene-d8	8.647	98	260950	48.294	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.079	95	100675	50.842	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031192.D
Acq On : 08 Sep 2022 05:41
Operator : JC/MD
Sample : N4446-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-04

Quant Time: Sep 08 13:58:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

