

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031194.D
 Acq On : 08 Sep 2022 06:28
 Operator : JC/MD
 Sample : N4446-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-WATER-LIQUID

Quant Time: Sep 08 13:58:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	70766	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	132252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49609	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108246	64.608	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 129.220%#	
35) Dibromofluoromethane	5.391	113	75437	52.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.680%	
50) Toluene-d8	8.652	98	259099	49.483	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.960%	
62) 4-Bromofluorobenzene	11.079	95	101881	53.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.180%	
Target Compounds						
					Qvalue	
16) Acetone	2.391	43	3932	6.778	ug/l	91
25) 2-Butanone	4.568	43	2582	3.005	ug/l	96
52) Toluene	8.726	92	1697	0.439	ug/l	86
67) Ethyl Benzene	10.195	91	5285	0.834	ug/l #	87
68) m/p-Xylenes	10.305	106	2351	1.001	ug/l	74
69) o-Xylene	10.646	106	3243	1.384	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031194.D
Acq On : 08 Sep 2022 06:28
Operator : JC/MD
Sample : N4446-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-WATER-LIQUID

Quant Time: Sep 08 13:58:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

