

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031197.D
 Acq On : 08 Sep 2022 07:38
 Operator : JC/MD
 Sample : PB147495ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147495ZHE#01

Quant Time: Sep 08 13:59:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72408	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	133899	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121606	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	113637	66.288	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	132.580	%#
35) Dibromofluoromethane	5.391	113	77498	53.112	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	106.220	%
50) Toluene-d8	8.653	98	268276	50.605	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.220	%
62) 4-Bromofluorobenzene	11.079	95	105881	54.500	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	109.000	%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	8861	14.928	ug/l	# 83
20) Methylene Chloride	2.794	84	3529	2.530	ug/l	# 76
37) Ethyl Acetate	4.715	43	10859m	5.602	ug/l	
43) Isopropyl Acetate	6.342	43	47374	15.500	ug/l	# 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031197.D
Acq On : 08 Sep 2022 07:38
Operator : JC/MD
Sample : PB147495ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147495ZHE#01

Quant Time: Sep 08 13:59:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

