

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
 Data File : VX031207.D
 Acq On : 08 Sep 2022 12:40
 Operator : JC/MD
 Sample : N4373-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 64 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE114D2-20220825RE

Quant Time: Sep 08 14:01:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 07 23:22:41 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63404	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	124643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106809	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45998	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91183	60.743	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.480%
35) Dibromofluoromethane	5.391	113	63101	46.457	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.920%
50) Toluene-d8	8.647	98	230000	46.607	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.220%
62) 4-Bromofluorobenzene	11.079	95	83488	46.165	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.320%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	640	0.628	ug/l	90
9) 1,1,2-Trichlorotrifluo...	2.331	101	10322	9.063	ug/l	97
27) cis-1,2-Dichloroethene	4.501	96	1243	0.913	ug/l	94
30) Chloroform	5.099	83	649	0.274	ug/l #	72
44) Trichloroethene	7.129	130	74009	47.647	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031207.D
Acq On : 08 Sep 2022 12:40
Operator : JC/MD
Sample : N4373-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 64 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE114D2-20220825RE

Quant Time: Sep 08 14:01:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 07 23:22:41 2022
Response via : Initial Calibration

