



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090722\
Data File : VX031150.D
Acq On : 07 Sep 2022 10:11
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X090722W.M
Title : SW846 8260
Last Update : Wed Sep 07 23:22:41 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.0	3935	PASS
75	95	30	60	54.0	11202	PASS
95	95	100	100	100.0	20737	PASS
96	95	5	9	6.7	1381	PASS
173	174	0.00	2	0.6	87	PASS
174	95	50	100	66.1	13715	PASS
175	174	5	9	7.5	1031	PASS
176	174	95	101	95.8	13133	PASS
177	176	5	9	6.5	860	PASS