

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037572.D
 Acq On : 07 Sep 2023 12:08
 Operator : JC/MD
 Sample : VX0907WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0907WBL01

Quant Time: Sep 08 01:11:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	134017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	222992	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198268	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84132	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79064	40.921	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.840%
35) Dibromofluoromethane	5.385	113	72760	49.268	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.540%
50) Toluene-d8	8.647	98	253428	47.308	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.620%
62) 4-Bromofluorobenzene	11.079	95	91895	43.651	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037572.D
Acq On : 07 Sep 2023 12:08
Operator : JC/MD
Sample : VX0907WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0907WBL01

Quant Time: Sep 08 01:11:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

