

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037578.D
 Acq On : 07 Sep 2023 14:43
 Operator : JC/MD
 Sample : 04287-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-090123

Quant Time: Sep 08 00:45:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246869	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87756	41.186	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.380%
35) Dibromofluoromethane	5.391	113	81112	49.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.220%
50) Toluene-d8	8.653	98	279484	47.126	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.260%
62) 4-Bromofluorobenzene	11.079	95	100924	43.303	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037578.D
Acq On : 07 Sep 2023 14:43
Operator : JC/MD
Sample : 04287-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-090123

Quant Time: Sep 08 00:45:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

