

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037582.D
 Acq On : 07 Sep 2023 16:39
 Operator : JC/MD
 Sample : 04311-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-02-133-090523-FD

Quant Time: Sep 08 00:46:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	155904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	261161	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93385	41.548	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.100%
35) Dibromofluoromethane	5.385	113	85513	49.441	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.880%
50) Toluene-d8	8.647	98	296518	47.262	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.520%
62) 4-Bromofluorobenzene	11.079	95	112194	45.504	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037582.D
Acq On : 07 Sep 2023 16:39
Operator : JC/MD
Sample : 04311-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-02-133-090523-FD

Quant Time: Sep 08 00:46:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

