

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037584.D
 Acq On : 07 Sep 2023 17:28
 Operator : JC/MD
 Sample : 04311-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-05-8-090523

Quant Time: Sep 08 00:47:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	141940	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	239347	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	214475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86402	42.223	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.440%
35) Dibromofluoromethane	5.391	113	77765	49.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.120%
50) Toluene-d8	8.647	98	273291	47.530	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.060%
62) 4-Bromofluorobenzene	11.079	95	98800	43.724	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.440%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	3.123	73	13765	2.484	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	5612	2.866	ug/l	78
44) Trichloroethene	7.129	130	820	0.448	ug/l	71

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037584.D
Acq On : 07 Sep 2023 17:28
Operator : JC/MD
Sample : 04311-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-05-8-090523

Quant Time: Sep 08 00:47:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

