

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037585.D
 Acq On : 07 Sep 2023 17:53
 Operator : JC/MD
 Sample : 04326-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OWBR-05-135-090623

Quant Time: Sep 08 00:48:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	248094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91557	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86680	40.540	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	81.080%
35) Dibromofluoromethane	5.391	113	78856	47.993	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.980%
50) Toluene-d8	8.653	98	279501	46.896	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.800%
62) 4-Bromofluorobenzene	11.079	95	101158	43.189	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.380%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	193846	126.971	ug/l #	79
16) Acetone	2.385	43	3021	3.612	ug/l #	79
21) trans-1,2-Dichloroethene	3.093	96	23072	13.294	ug/l	85
24) 1,1-Dichloroethane	3.611	63	22516	7.405	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	184200	90.030	ug/l	77
40) Benzene	6.049	78	2664	0.398	ug/l	99
42) 1,2-Dichloroethane	6.098	62	13009	4.955	ug/l	94
44) Trichloroethene	7.128	130	39939	21.038	ug/l	98
52) Toluene	8.726	92	7536	1.704	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037585.D
Acq On : 07 Sep 2023 17:53
Operator : JC/MD
Sample : 04326-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OWBR-05-135-090623

Quant Time: Sep 08 00:48:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

