

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037588.D
 Acq On : 07 Sep 2023 19:07
 Operator : JC/MD
 Sample : PB155358TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB155358TB

Quant Time: Sep 08 00:49:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	142862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246076	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	228397	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99214	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87639	42.551	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	85.100%		
35) Dibromofluoromethane	5.391	113	79414	48.729	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.460%		
50) Toluene-d8	8.647	98	287695	48.667	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.340%		
62) 4-Bromofluorobenzene	11.079	95	108632	46.761	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.520%		
Target Compounds						
16) Acetone	2.386	43	67606	83.905	ug/l #	83
20) Methylene Chloride	2.788	84	7439	4.128	ug/l #	81
25) 2-Butanone	4.568	43	16543	13.636	ug/l	95
29) Tetrahydrofuran	5.032	42	4272	5.522	ug/l #	82
37) Ethyl Acetate	4.733	43	2590	1.095	ug/l #	93
43) Isopropyl Acetate	6.348	43	94842	25.258	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037588.D
Acq On : 07 Sep 2023 19:07
Operator : JC/MD
Sample : PB155358TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB155358TB

Quant Time: Sep 08 00:49:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

