

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037591.D
 Acq On : 07 Sep 2023 20:22
 Operator : JC/MD
 Sample : 04327-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-03-127-090623

Quant Time: Sep 08 00:50:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	147526	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220564	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91239	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87878	41.318	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	82.640%	
35) Dibromofluoromethane	5.391	113	80394	49.249	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.500%	
50) Toluene-d8	8.647	98	280439	47.361	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	94.720%	
62) 4-Bromofluorobenzene	11.079	95	102907	44.223	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	88.440%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	65506	43.135	ug/l #	79
16) Acetone	2.386	43	13958	16.776	ug/l #	87
21) trans-1,2-Dichloroethene	3.099	96	4849	2.809	ug/l #	80
24) 1,1-Dichloroethane	3.611	63	13087	4.327	ug/l #	96
25) 2-Butanone	4.574	43	6330	5.053	ug/l #	89
27) cis-1,2-Dichloroethene	4.489	96	64606	31.745	ug/l	77
42) 1,2-Dichloroethane	6.098	62	8124	3.115	ug/l	92
44) Trichloroethene	7.129	130	222505	117.973	ug/l	94
64) Tetrachloroethene	9.275	164	512	0.318	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037591.D
Acq On : 07 Sep 2023 20:22
Operator : JC/MD
Sample : 04327-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-03-127-090623

Quant Time: Sep 08 00:50:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

