

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037594.D
 Acq On : 07 Sep 2023 21:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 08 00:52:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.581	0.528	9.1	70	0.00
3 P	Chloromethane	0.508	0.510	-0.4	72	0.00
4 C	Vinyl Chloride	0.553	0.571	-3.3#	79	0.00
5 T	Bromomethane	0.276	0.243	12.0	69	0.00
6 T	Chloroethane	0.325	0.313	3.7	77	0.00
7 T	Trichlorofluoromethane	0.906	0.896	1.1	78	0.00
8 T	Diethyl Ether	0.351	0.392	-11.7	78	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.555	-4.9	85	0.00
10 T	Methyl Iodide	0.642	0.723	-12.6	78	0.00
11 T	Tert butyl alcohol	0.169	0.185	-9.5	79	0.00
12 CM	1,1-Dichloroethene	0.515	0.557	-8.2#	84	0.00
13 T	Acrolein	0.129	0.143	-10.9	81	0.00
14 T	Allyl chloride	0.797	0.793	0.5	70	0.00
15 T	Acrylonitrile	0.302	0.390	-29.1#	91	0.00
16 T	Acetone	0.282	0.314	-11.3	82	0.00
17 T	Carbon Disulfide	1.309	1.068	18.4	64	0.00
18 T	Methyl Acetate	1.062	1.278	-20.3	85	0.00
19 T	Methyl tert-butyl Ether	1.952	2.190	-12.2	79	0.00
20 T	Methylene Chloride	0.631	0.688	-9.0	84	0.00
21 T	trans-1,2-Dichloroethene	0.585	0.599	-2.4	76	0.00
22 T	Diisopropyl ether	1.635	1.829	-11.9	79	0.00
23 T	Vinyl Acetate	1.253	1.354	-8.1	74	0.00
24 P	1,1-Dichloroethane	1.025	1.147	-11.9	81	0.00
25 T	2-Butanone	0.425	0.502	-18.1	83	0.00
26 T	2,2-Dichloropropane	0.916	0.518	43.4#	43#	0.00
27 T	cis-1,2-Dichloroethene	0.690	0.778	-12.8	81	0.00
28 T	Bromochloromethane	0.458	0.507	-10.7	78	0.00
29 T	Tetrahydrofuran	0.271	0.318	-17.3	82	0.00
30 C	Chloroform	1.151	1.273	-10.6#	80	0.00
31 T	Cyclohexane	0.811	0.801	1.2	82	0.00
32 T	1,1,1-Trichloroethane	1.031	1.093	-6.0	79	0.00
33 S	1,2-Dichloroethane-d4	0.721	0.731	-1.4	72	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.331	0.389	-17.5	85	0.00
36 T	1,1-Dichloropropene	0.470	0.461	1.9	77	0.00
37 T	Ethyl Acetate	0.481	0.535	-11.2	79	0.00
38 T	Carbon Tetrachloride	0.511	0.526	-2.9	79	0.00
39 T	Methylcyclohexane	0.494	0.534	-8.1	94	0.00
40 TM	Benzene	1.349	1.507	-11.7	81	0.00
41 T	Methacrylonitrile	0.250	0.290	-16.0	80	0.00
42 TM	1,2-Dichloroethane	0.529	0.527	0.4	70	0.00
43 T	Isopropyl Acetate	0.763	0.844	-10.6	76	0.00
44 TM	Trichloroethene	0.383	0.404	-5.5	79	0.00
45 C	1,2-Dichloropropane	0.333	0.391	-17.4#	83	0.00
46 T	Dibromomethane	0.262	0.292	-11.5	78	0.00
47 T	Bromodichloromethane	0.502	0.563	-12.2	78	0.00
48 T	Methyl methacrylate	0.372	0.411	-10.5	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
 Data File : VX037594.D
 Acq On : 07 Sep 2023 21:35
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 08 00:52:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.012	-20.0	85	0.00
50 S	Toluene-d8	1.201	1.319	-9.8	80	0.00
51 T	4-Methyl-2-Pentanone	0.463	0.558	-20.5	85	0.00
52 CM	Toluene	0.891	0.985	-10.5#	81	0.00
53 T	t-1,3-Dichloropropene	0.535	0.542	-1.3	68	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.602	-4.2	72	0.00
55 T	1,1,2-Trichloroethane	0.363	0.454	-25.1#	89	0.00
56 T	Ethyl methacrylate	0.569	0.687	-20.7	84	0.00
57 T	1,3-Dichloropropane	0.612	0.722	-18.0	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.299	-9.5	77	0.00
59 T	2-Hexanone	0.353	0.423	-19.8	84	0.00
60 T	Dibromochloromethane	0.386	0.467	-21.0	81	0.00
61 T	1,2-Dibromoethane	0.397	0.466	-17.4	83	0.00
62 S	4-Bromofluorobenzene	0.472	0.533	-12.9	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
64 T	Tetrachloroethene	0.365	0.363	0.5	80	0.00
65 PM	Chlorobenzene	1.057	1.152	-9.0	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.453	-16.2	85	0.00
67 C	Ethyl Benzene	1.860	1.985	-6.7#	83	0.00
68 T	m/p-Xylenes	0.718	0.778	-8.4	84	0.00
69 T	o-Xylene	0.708	0.784	-10.7	84	0.00
70 T	Styrene	1.167	1.322	-13.3	84	0.00
71 P	Bromoform	0.292	0.355	-21.6	82	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
73 T	Isopropylbenzene	3.634	4.067	-11.9	90	0.00
74 T	N-amyl acetate	1.412	1.519	-7.6	77	0.00
75 P	1,1,2,2-Tetrachloroethane	1.224	1.484	-21.2	91	0.00
76 T	1,2,3-Trichloropropane	1.065	1.182	-11.0	85	0.00
77 T	Bromobenzene	0.888	1.005	-13.2	86	0.00
78 T	n-propylbenzene	4.047	4.503	-11.3	90	0.00
79 T	2-Chlorotoluene	2.586	2.823	-9.2	86	0.00
80 T	1,3,5-Trimethylbenzene	3.032	3.437	-13.4	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.367	3.7	69	0.00
82 T	4-Chlorotoluene	2.985	3.157	-5.8	82	0.00
83 T	tert-Butylbenzene	2.891	3.557	-23.0	100	0.00
84 T	1,2,4-Trimethylbenzene	3.064	3.472	-13.3	89	0.00
85 T	sec-Butylbenzene	3.377	4.268	-26.4#	105	0.00
86 T	p-Isopropyltoluene	2.907	3.570	-22.8	101	0.00
87 T	1,3-Dichlorobenzene	1.662	1.812	-9.0	85	0.00
88 T	1,4-Dichlorobenzene	1.717	1.810	-5.4	83	0.00
89 T	n-Butylbenzene	2.410	2.867	-19.0	99	0.00
90 T	Hexachloroethane	0.485	0.640	-32.0#	102	0.00
91 T	1,2-Dichlorobenzene	1.627	1.848	-13.6	85	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.331	-16.5	81	0.00
93 T	1,2,4-Trichlorobenzene	0.952	1.099	-15.4	91	0.00
94 T	Hexachlorobutadiene	0.368	0.472	-28.3#	112	0.00
95 T	Naphthalene	4.171	4.202	-0.7	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090723\
Data File : VX037594.D
Acq On : 07 Sep 2023 21:35
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 08 00:52:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.946	1.155	-22.1	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6