

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090819\
 Data File : VX012220.D
 Acq On : 08 Sep 2019 13:51
 Operator : SY/SP
 Sample : K4622-04
 Misc : 6.57g/5.0mL/MSVOA X/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-11-A3-(0)

Quant Time: Sep 09 02:19:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	144619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	250645	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	218987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98768	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	99023	49.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.24%	
35) Dibromofluoromethane	5.48	113	75442	48.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.88%	
50) Toluene-d8	8.71	98	295111	52.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.34%	
62) 4-Bromofluorobenzene	11.13	95	117800	44.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.56%	
Target Compounds						
16) Acetone	2.43	43	22906	48.025	ug/l	97
29) Tetrahydrofuran	5.12	42	2829	7.734	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090819\
Data File : VX012220.D
Acq On : 08 Sep 2019 13:51
Operator : SY/SP
Sample : K4622-04
Misc : 6.57g/5.0mL/MSVOA X/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-11-A3-(0)

Quant Time: Sep 09 02:19:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

