

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX090819\
 Data File : VX012227.D
 Acq On : 08 Sep 2019 17:08
 Operator : SY/SP
 Sample : K4622-12
 Misc : 5.40g/5.0mL/MSVOA X/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-11-B2-(0)

Quant Time: Sep 09 03:17:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	143821	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	242783	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	212765	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	94493	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	95381	48.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.12%	
35) Dibromofluoromethane	5.48	113	73029	48.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.82%	
50) Toluene-d8	8.71	98	288697	52.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.38%	
62) 4-Bromofluorobenzene	11.13	95	113370	44.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.98%	
Target Compounds						
16) Acetone	2.43	43	11010	23.212	ug/l	99
29) Tetrahydrofuran	5.13	42	3421	9.404	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090819\
Data File : VX012227.D
Acq On : 08 Sep 2019 17:08
Operator : SY/SP
Sample : K4622-12
Misc : 5.40g/5.0mL/MSVOA X/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-11-B2-(0)

Quant Time: Sep 09 03:17:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

