

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090819\
 Data File : VX012235.D
 Acq On : 08 Sep 2019 20:52
 Operator : SY/SP
 Sample : K4622-03
 Misc : 4.87g/5.0mL/MSVOA X/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAR-11-A2-(0)

Quant Time: Sep 09 03:36:23 2019
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 03 15:26:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	145539	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	252921	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	222734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	95774	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	99392	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
35) Dibromofluoromethane	5.48	113	73889	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	8.71	98	300509	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
62) 4-Bromofluorobenzene	11.14	95	118696	44.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.44%	
Target Compounds						
16) Acetone	2.43	43	6415	13.365	ug/l	99
29) Tetrahydrofuran	5.13	42	3731	10.135	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX090819\
Data File : VX012235.D
Acq On : 08 Sep 2019 20:52
Operator : SY/SP
Sample : K4622-03
Misc : 4.87g/5.0mL/MSVOA X/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAR-11-A2-(0)

Quant Time: Sep 09 03:36:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 03 15:26:48 2019
Response via : Initial Calibration

