


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX090819\
Data File : VX012211.D
Acq On : 08 Sep 2019 09:33
Operator : SY/SP
Sample : BFB
Misc : 5.00µ/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X090319S.M
Title : SW846 8260
Last Update : Tue Sep 03 15:26:48 2019



AutoFind: Scans 1680, 1681, 1682; Background Corrected with Scan 1672

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	27901	PASS
75	95	30	60	50.9	76053	PASS
95	95	100	100	100.0	149312	PASS
96	95	5	9	6.4	9554	PASS
173	174	0.00	2	1.5	1485	PASS
174	95	50	100	64.7	96552	PASS
175	174	5	9	7.8	7485	PASS
176	174	95	101	101.0	97514	PASS
177	176	5	9	6.2	6078	PASS