

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018271.D
 Acq On : 08 Sep 2020 13:40
 Operator : JC/SP
 Sample : L3808-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Sep 09 07:06:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	450124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	715041	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	697331	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	340608	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	318016	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
35) Dibromofluoromethane	5.46	113	248060	50.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.88%	
50) Toluene-d8	8.70	98	879061	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	11.12	95	333418	46.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.44%	
Target Compounds						
16) Acetone	2.42	43	16088	6.607	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018271.D
Acq On : 08 Sep 2020 13:40
Operator : JC/SP
Sample : L3808-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Sep 09 07:06:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

