

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018277.D
 Acq On : 08 Sep 2020 16:03
 Operator : JC/SP
 Sample : L3940-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6

Quant Time: Sep 09 08:15:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	466026	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	736479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	709824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	377722	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	320626	49.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.12%	
35) Dibromofluoromethane	5.46	113	251185	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
50) Toluene-d8	8.70	98	908480	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	11.12	95	341135	46.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.82%	
Target Compounds						
16) Acetone	2.42	43	17606	6.984	ug/l #	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018277.D
Acq On : 08 Sep 2020 16:03
Operator : JC/SP
Sample : L3940-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6

Quant Time: Sep 09 08:15:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

