

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018281.D
 Acq On : 08 Sep 2020 17:33
 Operator : JC/SP
 Sample : PB131451ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131451ZHE#02

Quant Time: Sep 09 02:23:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	434903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	703318	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	709313	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	352126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	310668	51.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.92%	
35) Dibromofluoromethane	5.46	113	248070	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
50) Toluene-d8	8.70	98	903719	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
62) 4-Bromofluorobenzene	11.12	95	336146	47.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.78%	
Target Compounds						
16) Acetone	2.42	43	59192	25.160	ug/l	97
18) Methyl Acetate	2.75	43	5977	1.053	ug/l #	66
20) Methylene Chloride	2.84	84	17599	2.550	ug/l	97
43) Isopropyl Acetate	6.42	43	105355	8.657	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018281.D
Acq On : 08 Sep 2020 17:33
Operator : JC/SP
Sample : PB131451ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131451ZHE#02

Quant Time: Sep 09 02:23:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

