

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018282.D
 Acq On : 08 Sep 2020 17:56
 Operator : JC/SP
 Sample : PB131451ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131451ZHE#03

Quant Time: Sep 09 02:24:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	458952	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	738885	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	744028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	361543	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	330781	51.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.84%	
35) Dibromofluoromethane	5.46	113	254337	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
50) Toluene-d8	8.70	98	943738	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	11.12	95	353521	47.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.88%	
Target Compounds						
16) Acetone	2.42	43	58847	23.703	ug/l	93
18) Methyl Acetate	2.76	43	7682	1.282	ug/l #	86
20) Methylene Chloride	2.84	84	17651	2.364	ug/l	94
43) Isopropyl Acetate	6.42	43	105115	8.222	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018282.D
Acq On : 08 Sep 2020 17:56
Operator : JC/SP
Sample : PB131451ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131451ZHE#03

Quant Time: Sep 09 02:24:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

