

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018284.D
 Acq On : 08 Sep 2020 18:42
 Operator : JC/SP
 Sample : PB131451ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB131451ZHE#05

Quant Time: Sep 09 02:27:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	441122	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	706250	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	703255	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	345276	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	314924	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
35) Dibromofluoromethane	5.47	113	245295	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
50) Toluene-d8	8.70	98	903763	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	11.12	95	343326	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
Target Compounds						
16) Acetone	2.42	43	59376	24.883	ug/l	90
18) Methyl Acetate	2.75	43	6858	1.191	ug/l	100
20) Methylene Chloride	2.84	84	18933	2.777	ug/l	90
43) Isopropyl Acetate	6.42	43	104895	8.584	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018284.D
Acq On : 08 Sep 2020 18:42
Operator : JC/SP
Sample : PB131451ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB131451ZHE#05

Quant Time: Sep 09 02:27:41 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

