

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_X\DATA\VX090820\  
 Data File : VX018298.D  
 Acq On : 09 Sep 2020 01:08  
 Operator : JC/SP  
 Sample : L3877-08  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 40 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 TP-2

Quant Time: Sep 09 03:11:12 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X082120W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Aug 21 03:03:56 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 5.64  | 168  | 457305   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.83  | 114  | 734879   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 10.10 | 117  | 712033   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.06 | 152  | 360360   | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 6.03  | 65   | 325338   | 51.25  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 102.50% |          |
| 35) Dibromofluoromethane    | 5.47  | 113  | 250004   | 49.95  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.90%  |          |
| 50) Toluene-d8              | 8.70  | 98   | 927521   | 49.76  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.52%  |          |
| 62) 4-Bromofluorobenzene    | 11.12 | 95   | 355401   | 48.46  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 96.92%  |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 2.42  | 43   | 44927    | 18.161 | ug/l    | 99       |
| 18) Methyl Acetate          | 2.75  | 43   | 6907     | 1.157  | ug/l    | 97       |
| 20) Methylene Chloride      | 2.84  | 84   | 20818    | 3.018  | ug/l    | 92       |
| 40) Benzene                 | 6.12  | 78   | 102208   | 5.132  | ug/l    | 99       |
| 43) Isopropyl Acetate       | 6.41  | 43   | 104536   | 8.221  | ug/l    | 99       |
| 95) Naphthalene             | 13.82 | 128  | 149643   | 6.279  | ug/l    | 96       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_X\DATA\VX090820\  
Data File : VX018298.D  
Acq On : 09 Sep 2020 01:08  
Operator : JC/SP  
Sample : L3877-08  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 40 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
TP-2

Quant Time: Sep 09 03:11:12 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X082120W.M  
Quant Title : SW846 8260  
QLast Update : Fri Aug 21 03:03:56 2020  
Response via : Initial Calibration

