

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018301.D
 Acq On : 09 Sep 2020 02:16
 Operator : JC/SP
 Sample : L3808-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B44-090320-DP

Quant Time: Sep 09 03:16:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	433269	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	705262	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	693038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	332921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	314023	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	5.47	113	242170	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
50) Toluene-d8	8.70	98	881917	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
62) 4-Bromofluorobenzene	11.12	95	333595	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
Target Compounds						
16) Acetone	2.42	43	22736	9.701	ug/l	100
20) Methylene Chloride	2.84	84	14912	1.990	ug/l	90
43) Isopropyl Acetate	6.42	43	96702	7.925	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018301.D
Acq On : 09 Sep 2020 02:16
Operator : JC/SP
Sample : L3808-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B44-090320-DP

Quant Time: Sep 09 03:16:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

