

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018303.D
 Acq On : 09 Sep 2020 03:02
 Operator : JC/SP
 Sample : L3915-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-PM-A1-0-4-D

Quant Time: Sep 09 03:52:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	422374	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	696924	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	682786	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	330566	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	312876	53.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.72%	
35) Dibromofluoromethane	5.46	113	243410	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
50) Toluene-d8	8.70	98	875661	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.12	95	324688	46.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.36%	
Target Compounds						
16) Acetone	2.42	43	62851	27.508	ug/l	97
18) Methyl Acetate	2.75	43	6336	1.149	ug/l #	80
20) Methylene Chloride	2.84	84	17645	2.671	ug/l	87
43) Isopropyl Acetate	6.41	43	69821	5.790	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018303.D
Acq On : 09 Sep 2020 03:02
Operator : JC/SP
Sample : L3915-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
WC-PM-A1-0-4-D

Quant Time: Sep 09 03:52:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

