

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018307.D
 Acq On : 09 Sep 2020 04:32
 Operator : JC/SP
 Sample : L3926-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BR-1

Quant Time: Sep 09 05:02:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	459004	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	718093	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	717223	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	368030	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	324066	50.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.72%	
35) Dibromofluoromethane	5.46	113	252555	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	8.70	98	919624	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
62) 4-Bromofluorobenzene	11.12	95	343420	47.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.84%	
Target Compounds						
16) Acetone	2.42	43	67832	27.319	ug/l	89
20) Methylene Chloride	2.84	84	15697	1.970	ug/l	96
43) Isopropyl Acetate	6.42	43	42539	3.424	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018307.D
Acq On : 09 Sep 2020 04:32
Operator : JC/SP
Sample : L3926-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BR-1

Quant Time: Sep 09 05:02:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

