

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018269.D
 Acq On : 08 Sep 2020 12:55
 Operator : JC/SP
 Sample : L3925-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 09 07:03:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	451315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	711630	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	682281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	342756	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	318361	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
35) Dibromofluoromethane	5.47	113	248198	51.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.42%	
50) Toluene-d8	8.70	98	862110	47.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.52%	
62) 4-Bromofluorobenzene	11.12	95	327881	46.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.34%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018269.D
Acq On : 08 Sep 2020 12:55
Operator : JC/SP
Sample : L3925-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 09 07:03:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX018269.D

Acq: 08 Sep 2020 12:55

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tgt Ion:168 Resp: 451315

Ion Ratio Lower Upper




168 100

99 62.3 46.3 69.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.812 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX018269.D


Acq: 08 Sep 2020 12:55

Tgt Ion: 65 Resp: 318361

Ion Ratio Lower Upper



65 100


67 50.3 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.84 min Scan# 943

Delta R.T. -0.00 min

Lab File: VX018269.D

Acq: 08 Sep 2020 12:55

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 711630

Ion Ratio Lower Upper

114 100

63 20.2 0.0 42.6




88 16.7 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 51.211 ug/l

RT: 5.47 min Scan# 719

Delta R.T. -0.00 min

Lab File: VX018269.D

Acq: 08 Sep 2020 12:55


Tgt Ion:113 Resp: 248198

Ion Ratio Lower Upper

113 100

111 102.6 81.6 122.4



192 19.6 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.759 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX018269.D

Acq: 08 Sep 2020 12:55

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-WATER-REF-4

Tgt Ion: 98 Resp: 862110

Ion Ratio Lower Upper




98 100

100 65.2 52.2 78.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V



#62

4-Bromofluorobenzene

Concen: 46.167 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX018269.D

Acq: 08 Sep 2020 12:55


Tgt Ion: 95 Resp: 327881

Ion Ratio Lower Upper

95 100

174 80.4 0.0 160.2



176 77.4 0.0 155.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX018269.D
Acq: 08 Sep 2020 12:55

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 682281
Ion Ratio Lower Upper
117 100
82 55.0 44.2 66.2
119 31.7 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX018269.D
Acq: 08 Sep 2020 12:55

Tgt Ion:152 Resp: 342756
Ion Ratio Lower Upper
152 100
115 60.1 41.5 124.5
150 156.6 0.0 343.8

