

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
 Data File : VX018270.D
 Acq On : 08 Sep 2020 13:18
 Operator : JC/SP
 Sample : L3925-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 09 07:04:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	456024	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	721458	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	681210	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	352407	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	320865	50.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.36%	
35) Dibromofluoromethane	5.46	113	250094	50.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.80%	
50) Toluene-d8	8.70	98	870850	47.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.18%	
62) 4-Bromofluorobenzene	11.12	95	329652	45.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.56%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX090820\
Data File : VX018270.D
Acq On : 08 Sep 2020 13:18
Operator : JC/SP
Sample : L3925-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Sep 09 07:04:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX018270.D

Acq: 08 Sep 2020 13:18

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tgt Ion: 168 Resp: 456024

Ion Ratio Lower Upper




168 100

99 59.6 46.3 69.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.683 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX018270.D


Acq: 08 Sep 2020 13:18

Tgt Ion: 65 Resp: 320865

Ion Ratio Lower Upper



65 100


67 49.8 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.01 min

Lab File: VX018270.D

Acq: 08 Sep 2020 13:18

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-SAMPLE-MANA


Tgt Ion:114 Resp: 721458

Ion Ratio Lower Upper

114 100

63 21.1 0.0 42.6




88 16.6 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 50.900 ug/l

RT: 5.46 min Scan# 718

Delta R.T. -0.01 min

Lab File: VX018270.D

Acq: 08 Sep 2020 13:18


Tgt Ion:113 Resp: 250094

Ion Ratio Lower Upper

113 100

111 101.0 81.6 122.4



192 19.4 15.8 23.8




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 47.586 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX018270.D

Acq: 08 Sep 2020 13:18

Instrument :

MSVOA_X

ClientSampleId :


STORAGE-BLANK-SAMPLE-MANA

Tgt Ion: 98 Resp: 870850

Ion Ratio Lower Upper




98 100

100 65.0 52.2 78.4



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): VX



#62

4-Bromofluorobenzene

Concen: 45.784 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX018270.D

Acq: 08 Sep 2020 13:18


Tgt Ion: 95 Resp: 329652

Ion Ratio Lower Upper

95 100

174 81.5 0.0 160.2



176 76.6 0.0 155.8




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX018270.D
Acq: 08 Sep 2020 13:18

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Tgt Ion:117 Resp: 681210
Ion Ratio Lower Upper
117 100
82 56.3 44.2 66.2
119 32.2 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX018270.D
Acq: 08 Sep 2020 13:18

Tgt Ion:152 Resp: 352407
Ion Ratio Lower Upper
152 100
115 56.4 41.5 124.5
150 150.0 0.0 343.8

