

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024128.D
 Acq On : 08 Sep 2021 11:11
 Operator : JC/MD
 Sample : VX0908WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS01

Quant Time: Sep 09 05:42:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	175794	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	265538	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	127290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	115525	47.085	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.180%
35) Dibromofluoromethane	5.391	113	87025	49.516	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.040%
50) Toluene-d8	8.653	98	308367	46.099	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.200%
62) 4-Bromofluorobenzene	11.085	95	124001	50.356	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.720%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	39061	21.060	ug/l	100
3) Chloromethane	1.294	50	35894	17.629	ug/l	100
4) Vinyl Chloride	1.374	62	35369	17.811	ug/l	99
5) Bromomethane	1.599	94	27685	22.175	ug/l	97
6) Chloroethane	1.685	64	22786	15.826	ug/l	96
7) Trichlorofluoromethane	1.886	101	62036	19.783	ug/l	96
8) Diethyl Ether	2.142	74	20184	18.431	ug/l	91
9) 1,1,2-Trichlorotrifluo...	2.331	101	35518	20.128	ug/l	95
10) Methyl Iodide	2.453	142	41052	16.708	ug/l	93
11) Tert butyl alcohol	2.971	59	50275	101.324	ug/l	99
12) 1,1-Dichloroethene	2.319	96	32544	19.290	ug/l	93
13) Acrolein	2.239	56	23695	80.249	ug/l	95
14) Allyl chloride	2.666	41	56113	18.048	ug/l	87
15) Acrylonitrile	3.069	53	100100	90.544	ug/l	97
16) Acetone	2.386	43	108330	98.748	ug/l	89
17) Carbon Disulfide	2.514	76	75386	18.735	ug/l	99
18) Methyl Acetate	2.709	43	60917	22.244	ug/l	# 89
19) Methyl tert-butyl Ether	3.123	73	117994	20.095	ug/l	98
20) Methylene Chloride	2.794	84	38506	18.376	ug/l	92
21) trans-1,2-Dichloroethene	3.093	96	35780	19.855	ug/l	95
22) Diisopropyl ether	3.770	45	115212	18.605	ug/l	97
23) Vinyl Acetate	3.727	43	500786	99.970	ug/l	# 94
24) 1,1-Dichloroethane	3.611	63	66077	19.173	ug/l	98
25) 2-Butanone	4.568	43	151156	90.969	ug/l	95
26) 2,2-Dichloropropane	4.483	77	59698	21.106	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	41076	19.156	ug/l	95
28) Bromochloromethane	4.904	49	24576	15.683	ug/l	97
29) Tetrahydrofuran	5.019	42	94430	87.737	ug/l	91
30) Chloroform	5.105	83	71310	19.542	ug/l	96
31) Cyclohexane	5.477	56	58898	18.685	ug/l	99
32) 1,1,1-Trichloroethane	5.391	97	64595	20.673	ug/l	98
36) 1,1-Dichloropropene	5.696	75	52378	21.300	ug/l	98
37) Ethyl Acetate	4.727	43	55801	18.460	ug/l	95
38) Carbon Tetrachloride	5.678	117	56780	22.143	ug/l	98
39) Methylcyclohexane	7.385	83	63968	21.364	ug/l	95
40) Benzene	6.044	78	148979	20.488	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024128.D
 Acq On : 08 Sep 2021 11:11
 Operator : JC/MD
 Sample : VX0908WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS01

Quant Time: Sep 09 05:42:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	31553	19.255	ug/l	91
42) 1,2-Dichloroethane	6.098	62	63557	22.240	ug/l	93
43) Isopropyl Acetate	6.348	43	91930	19.443	ug/l	93
44) Trichloroethene	7.135	130	41805	20.812	ug/l	96
45) 1,2-Dichloropropane	7.434	63	37735	19.543	ug/l	98
46) Dibromomethane	7.586	93	27957	20.806	ug/l	99
47) Bromodichloromethane	7.824	83	53205	21.888	ug/l	98
48) Methyl methacrylate	7.696	41	56040	24.296	ug/l	85
49) 1,4-Dioxane	7.659	88	19159	393.027	ug/l #	94
51) 4-Methyl-2-Pentanone	8.580	43	297645	97.386	ug/l	90
52) Toluene	8.720	92	96365	20.452	ug/l	96
53) t-1,3-Dichloropropene	8.982	75	57197	19.575	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	61912	21.020	ug/l #	89
55) 1,1,2-Trichloroethane	9.153	97	40989	21.130	ug/l	99
56) Ethyl methacrylate	9.122	69	46847	14.678	ug/l #	80
57) 1,3-Dichloropropane	9.311	76	65890	20.479	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	141101	90.332	ug/l	98
59) 2-Hexanone	9.433	43	236711	100.756	ug/l	86
60) Dibromochloromethane	9.525	129	41324	20.325	ug/l	100
61) 1,2-Dibromoethane	9.610	107	42381	20.681	ug/l	99
64) Tetrachloroethene	9.275	164	43306	21.983	ug/l	93
65) Chlorobenzene	10.079	112	105880	20.572	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.165	131	39894	21.778	ug/l	99
67) Ethyl Benzene	10.195	91	191698	21.329	ug/l	99
68) m/p-Xylenes	10.305	106	144575	42.232	ug/l	98
69) o-Xylene	10.646	106	69850	20.763	ug/l	100
70) Styrene	10.659	104	104872	19.135	ug/l	96
71) Bromoform	10.805	173	28322	20.223	ug/l #	98
73) Isopropylbenzene	10.963	105	192739	19.851	ug/l	99
74) N-amyl acetate	10.848	43	80375	18.318	ug/l	90
75) 1,1,2,2-Tetrachloroethane	11.213	83	62423	19.618	ug/l	97
76) 1,2,3-Trichloropropane	11.244	75	72089	24.278	ug/l	91
77) Bromobenzene	11.201	156	47537	19.879	ug/l	96
78) n-propylbenzene	11.305	91	216566	19.502	ug/l	99
79) 2-Chlorotoluene	11.366	91	132519	19.744	ug/l	99
80) 1,3,5-Trimethylbenzene	11.457	105	163804	20.530	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	18120	17.950	ug/l #	87
82) 4-Chlorotoluene	11.457	91	153784	20.186	ug/l	99
83) tert-Butylbenzene	11.719	119	157166	20.282	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	164913	20.729	ug/l	98
85) sec-Butylbenzene	11.896	105	187621	19.124	ug/l	100
86) p-Isopropyltoluene	12.012	119	169438	20.984	ug/l	97
87) 1,3-Dichlorobenzene	11.975	146	84255	19.474	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	85426	19.797	ug/l	97
89) n-Butylbenzene	12.335	91	147603	20.473	ug/l	97
90) Hexachloroethane	12.542	117	25957	19.367	ug/l	97
91) 1,2-Dichlorobenzene	12.335	146	81845	19.527	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	12444	18.120	ug/l	91
93) 1,2,4-Trichlorobenzene	13.591	180	50955	20.199	ug/l	99
94) Hexachlorobutadiene	13.725	225	25477	22.659	ug/l	98
95) Naphthalene	13.780	128	159514	17.976	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	50436	20.075	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
Data File : VX024128.D
Acq On : 08 Sep 2021 11:11
Operator : JC/MD
Sample : VX0908WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01

Quant Time: Sep 09 05:42:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX090821\
 Data File : VX024128.D
 Acq On : 08 Sep 2021 11:11
 Operator : JC/MD
 Sample : VX0908WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sample ID :
 VX0908WBS01

Quant Time: Sep 09 05:42:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:168 Resp: 175794
Ion Ratio Lower Upper
168 100
99 58.6 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 21.060 ug/l
RT: 1.173 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 85 Resp: 39061
Ion Ratio Lower Upper
85 100
87 32.7 16.4 49.2









#3
Chloromethane
Concen: 17.629 ug/l
RT: 1.294 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 50 Resp: 35894
Ion Ratio Lower Upper
50 100
52 32.9 26.4 39.6



#4
Vinyl Chloride
Concen: 17.811 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 62 Resp: 35369
Ion Ratio Lower Upper
62 100
64 32.6 25.7 38.5









#5
Bromomethane
Concen: 22.175 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 94 Resp: 27685
Ion Ratio Lower Upper
94 100
96 98.0 76.4 114.6



#6
Chloroethane
Concen: 15.826 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 64 Resp: 22786
Ion Ratio Lower Upper
64 100
66 33.8 25.4 38.0









#7
Trichlorofluoromethane
Concen: 19.783 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:101 Resp: 62036
Ion Ratio Lower Upper
101 100
103 69.7 53.2 79.8



#8
Diethyl Ether
Concen: 18.431 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 74 Resp: 20184
Ion Ratio Lower Upper
74 100
45 97.1 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.128 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:101 Resp: 35518
Ion Ratio Lower Upper
101 100
85 46.6 34.9 52.3
151 81.7 61.4 92.2



#10
Methyl Iodide
Concen: 16.708 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:142 Resp: 41052
Ion Ratio Lower Upper
142 100
127 47.4 33.3 49.9
141 14.1 11.4 17.0









#11
Tert butyl alcohol
Concen: 101.324 ug/l
RT: 2.971 min Scan# 309
Delta R.T. -0.012 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01


Tgt Ion: 59 Resp: 50275
Ion Ratio Lower Upper
59 100
57 9.5 7.8 11.8



#12
1,1-Dichloroethene
Concen: 19.290 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 96 Resp: 32544
Ion Ratio Lower Upper
96 100
61 166.6 124.6 187.0
98 60.3 50.6 76.0









#13
Acrolein
Concen: 80.249 ug/l
RT: 2.239 min Scan# 189
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 56 Resp: 23695
Ion Ratio Lower Upper
56 100
55 74.5 56.4 84.6



#14
Allyl chloride
Concen: 18.048 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 41 Resp: 56113
Ion Ratio Lower Upper
41 100
39 76.2 51.6 77.4
76 32.4 30.5 45.7









#15
Acrylonitrile
Concen: 90.544 ug/l
RT: 3.069 min Scan# 325
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 53 Resp: 100100
Ion Ratio Lower Upper
53 100
52 83.3 65.3 97.9
51 38.7 28.2 42.2



#16
Acetone
Concen: 98.748 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 43 Resp: 108330
Ion Ratio Lower Upper
43 100
58 28.6 27.8 41.8









#17
Carbon Disulfide
Concen: 18.735 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 76 Resp: 75386
Ion Ratio Lower Upper
76 100
78 9.3 7.2 10.8



#18
Methyl Acetate
Concen: 22.244 ug/l
RT: 2.709 min Scan# 266
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 43 Resp: 60917
Ion Ratio Lower Upper
43 100
74 21.9 21.9 32.9#









#19
Methyl tert-butyl Ether
Concen: 20.095 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 73 Resp: 117994
Ion Ratio Lower Upper
73 100
57 20.3 17.1 25.7



#20
Methylene Chloride
Concen: 18.376 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 84 Resp: 38506
Ion Ratio Lower Upper
84 100
49 122.8 88.6 132.8
51 38.2 27.0 40.6
86 64.3 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 19.855 ug/l
RT: 3.093 min Scan# 329
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 96 Resp: 35780
Ion Ratio Lower Upper
96 100
61 142.0 107.7 161.5
98 65.8 50.4 75.6



#22
Diisopropyl ether
Concen: 18.605 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 45 Resp: 115212
Ion Ratio Lower Upper
45 100
43 75.9 58.0 87.0
87 29.8 25.4 38.0
59 12.7 10.2 15.2









#23
 Vinyl Acetate
 Concen: 99.970 ug/l
 RT: 3.727 min Scan# 433
 Delta R.T. -0.006 min
 Lab File: VX024128.D
 Acq: 08 Sep 2021 11:11




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0908WBS01


Tgt Ion: 43 Resp: 500786
 Ion Ratio Lower Upper
 43 100
 86 9.9 9.9 14.9#



#24
 1,1-Dichloroethane
 Concen: 19.173 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. -0.006 min
 Lab File: VX024128.D
 Acq: 08 Sep 2021 11:11

Tgt Ion: 63 Resp: 66077
 Ion Ratio Lower Upper
 63 100
 98 7.1 3.8 11.2
 100 3.8 2.5 7.5









#25
2-Butanone
Concen: 90.969 ug/l
RT: 4.568 min Scan# 571
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01


Tgt Ion: 43 Resp: 151156
Ion Ratio Lower Upper
43 100
72 27.0 23.7 35.5



#26
2,2-Dichloropropane
Concen: 21.106 ug/l
RT: 4.483 min Scan# 557
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 77 Resp: 59698
Ion Ratio Lower Upper
77 100
97 22.4 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 19.156 ug/l

RT: 4.495 min Scan# 559





Delta R.T. -0.006 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion: 96 Resp: 41076

Ion	Ratio	Lower	Upper
96	100		
61	152.2	0.0	285.6
98	64.3	0.0	129.6



#28

Bromochloromethane

Concen: 15.683 ug/l

RT: 4.904 min Scan# 626




Delta R.T. -0.006 min


Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion: 49 Resp: 24576

Ion	Ratio	Lower	Upper
49	100		
129	2.5	0.0	4.0
130	82.8	68.2	102.2





#29

Tetrahydrofuran

Concen: 87.737 ug/l

RT: 5.019 min Scan# 645

Delta R.T. 0.000 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11





Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS01

Tgt Ion:	42	Resp:	94430
Ion	Ratio	Lower	Upper
42	100		
72	45.0	41.7	62.5
71	42.4	38.4	57.6



#30

Chloroform

Concen: 19.542 ug/l




RT: 5.105 min Scan# 659


Delta R.T. -0.000 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion:	83	Resp:	71310
Ion	Ratio	Lower	Upper
83	100		
85	68.0	52.0	78.0









#31
Cyclohexane
Concen: 18.685 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:	56	Resp:	58898
Ion Ratio	Lower	Upper	
56	100		
69	31.6	25.3	37.9
84	88.3	71.4	107.0



#32
1,1,1-Trichloroethane
Concen: 20.673 ug/l
RT: 5.391 min Scan# 706
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:	97	Resp:	64595
Ion Ratio	Lower	Upper	
97	100		
99	64.4	51.5	77.3
61	45.3	34.2	51.4









#33

1,2-Dichloroethane-d4
Concen: 47.085 ug/l
RT: 5.958 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01




Tgt Ion: 65 Resp: 115525
Ion Ratio Lower Upper
65 100
67 50.1 0.0 105.4




#34

1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 114 Resp: 265538
Ion Ratio Lower Upper
114 100
63 20.7 0.0 40.8
88 16.6 0.0 32.8





#35

Dibromofluoromethane

Concen: 49.516 ug/l

RT: 5.391 min Scan# 706

Delta R.T. -0.006 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion:113 Resp: 87025

Ion Ratio Lower Upper

113 100

111 103.5 81.9 122.9





192 19.4 16.7 25.1

Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS01



#36

1,1-Dichloropropene

Concen: 21.300 ug/l

RT: 5.696 min Scan# 756

Delta R.T. -0.006 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11




Tgt Ion: 75 Resp: 52378


Ion Ratio Lower Upper

75 100

110 35.2 18.1 54.4

77 31.0 25.3 37.9





#37

Ethyl Acetate

Concen: 18.460 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11





Instrument :

MSVOA_X

ClientSampleId :

VX0908WBS01

Tgt Ion:	43	Resp:	55801
Ion	Ratio	Lower	Upper
43	100		
61	14.1	12.8	19.2
70	11.0	10.5	15.7



#38

Carbon Tetrachloride

Concen: 22.143 ug/l




RT: 5.678 min Scan# 753


Delta R.T. -0.012 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion:	117	Resp:	56780
Ion	Ratio	Lower	Upper
117	100		
119	93.0	76.2	114.2
121	29.5	24.8	37.2









#39
Methylcyclohexane
Concen: 21.364 ug/l
RT: 7.385 min Scan# 1033
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 83 Resp: 63968
Ion Ratio Lower Upper
83 100
55 75.6 58.9 88.3
98 41.9 38.1 57.1



#40
Benzene
Concen: 20.488 ug/l
RT: 6.044 min Scan# 813
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 78 Resp: 148979
Ion Ratio Lower Upper
78 100
77 24.1 18.9 28.3









#41
Methacrylonitrile
Concen: 19.255 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01

Tgt Ion: 41 Resp: 31553




Ion	Ratio	Lower	Upper
41	100		
39	56.7	42.5	63.7
67	68.0	65.0	97.6
52	30.9	25.3	37.9




#42
1,2-Dichloroethane
Concen: 22.240 ug/l
RT: 6.098 min Scan# 822
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 62 Resp: 63557

Ion	Ratio	Lower	Upper
62	100		
98	7.4	0.0	20.2









#43
Isopropyl Acetate
Concen: 19.443 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:	43	Resp:	91930
Ion	Ratio	Lower	Upper
43	100		
61	18.9	18.2	27.4
87	13.3	12.5	18.7



#44
Trichloroethene
Concen: 20.812 ug/l
RT: 7.135 min Scan# 992
Delta R.T. 0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:	130	Resp:	41805
Ion	Ratio	Lower	Upper
130	100		
95	95.5	0.0	198.0









#45

1,2-Dichloropropane
Concen: 19.543 ug/l
RT: 7.434 min Scan# 1041
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01




Tgt Ion: 63 Resp: 37735
Ion Ratio Lower Upper
63 100
65 33.0 25.7 38.5




#46

Dibromomethane
Concen: 20.806 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 93 Resp: 27957
Ion Ratio Lower Upper
93 100
95 83.9 66.8 100.2
174 101.7 83.1 124.7









#47
Bromodichloromethane
Concen: 21.888 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 83 Resp: 53205
Ion Ratio Lower Upper
83 100
85 62.6 51.4 77.0
127 9.4 6.8 10.2



#48
Methyl methacrylate
Concen: 24.296 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 41 Resp: 56040
Ion Ratio Lower Upper
41 100
69 77.9 76.0 114.0
39 62.8 44.6 67.0









#49
1,4-Dioxane
Concen: 393.027 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01


Tgt Ion: 88 Resp: 19159
Ion Ratio Lower Upper
88 100
43 36.6 23.2 34.8#
58 67.0 55.0 82.4



#50
Toluene-d8
Concen: 46.099 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 98 Resp: 308367
Ion Ratio Lower Upper
98 100
100 64.0 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 97.386 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 43 Resp: 297645
Ion Ratio Lower Upper
43 100
58 37.6 35.4 53.0



#52
Toluene
Concen: 20.452 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 92 Resp: 96365
Ion Ratio Lower Upper
92 100
91 177.5 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 19.575 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01


Tgt Ion: 75 Resp: 57197
Ion Ratio Lower Upper
75 100
77 30.0 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 21.020 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 75 Resp: 61912
Ion Ratio Lower Upper
75 100
77 31.3 25.7 38.5
39 55.0 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 21.130 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:	97	Resp:	40989
Ion	Ratio	Lower	Upper
97	100		
83	86.4	69.3	103.9
85	57.0	44.3	66.5
99	63.9	50.3	75.5



#56
Ethyl methacrylate
Concen: 14.678 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:	69	Resp:	46847
Ion	Ratio	Lower	Upper
69	100		
41	71.3	46.8	70.2#
39	44.2	24.7	37.1#









#57

1,3-Dichloropropane
Concen: 20.479 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01




Tgt Ion: 76 Resp: 65890
Ion Ratio Lower Upper
76 100
78 33.1 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 90.332 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 63 Resp: 141101
Ion Ratio Lower Upper
63 100
106 28.3 23.5 35.3









#59
2-Hexanone
Concen: 100.756 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 43 Resp: 236711
Ion Ratio Lower Upper
43 100
58 51.4 30.9 92.8



#60
Dibromochloromethane
Concen: 20.325 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 129 Resp: 41324
Ion Ratio Lower Upper
129 100
127 76.6 38.4 115.2









#61
1,2-Dibromoethane
Concen: 20.681 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:107 Resp: 42381
Ion Ratio Lower Upper
107 100
109 95.2 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 50.356 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 95 Resp: 124001
Ion Ratio Lower Upper
95 100
174 80.4 0.0 154.6
176 78.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:117 Resp: 246646
Ion Ratio Lower Upper
117 100
82 58.1 47.4 71.2
119 32.2 26.6 39.8



#64
Tetrachloroethene
Concen: 21.983 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:164 Resp: 43306
Ion Ratio Lower Upper
164 100
166 124.7 103.4 155.2
129 98.7 70.8 106.2
131 97.2 72.5 108.7









#65
Chlorobenzene
Concen: 20.572 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.007 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:112 Resp: 105880
Ion Ratio Lower Upper
112 100
114 33.7 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 21.778 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:131 Resp: 39894
Ion Ratio Lower Upper
131 100
133 95.0 47.7 143.1
119 65.5 32.1 96.5









#67
Ethyl Benzene
Concen: 21.329 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 91 Resp: 191698
Ion Ratio Lower Upper
91 100
106 30.1 24.6 37.0



#68
m/p-Xylenes
Concen: 42.232 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 106 Resp: 144575
Ion Ratio Lower Upper
106 100
91 207.2 162.8 244.2









#69
o-Xylene
Concen: 20.763 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:106 Resp: 69850
Ion Ratio Lower Upper
106 100
91 216.8 108.1 324.2



#70
Styrene
Concen: 19.135 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:104 Resp: 104872
Ion Ratio Lower Upper
104 100
78 54.2 40.0 60.0
103 55.9 43.4 65.0





#71

Bromoform

Concen: 20.223 ug/l

RT: 10.805 min Scan# 1594





Delta R.T. -0.000 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion:173	Resp:	28322
Ion Ratio	Lower	Upper
173	100	
175	49.3	23.9 71.8
254	0.0	0.1 0.1#

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX024128.D

Acq: 08 Sep 2021 11:11

Tgt Ion:152	Resp:	127290
Ion Ratio	Lower	Upper
152	100	
115	70.4	41.1 123.3
150	162.0	0.0 349.0









#73
Isopropylbenzene
Concen: 19.851 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:105 Resp: 192739
Ion Ratio Lower Upper
105 100
120 25.8 13.2 39.6



#74
N-ethyl acetate
Concen: 18.318 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 43 Resp: 80375
Ion Ratio Lower Upper
43 100
70 39.7 39.0 58.6
55 24.4 21.5 32.3
61 22.2 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 19.618 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 83 Resp: 62423
Ion Ratio Lower Upper
83 100
131 11.2 5.1 15.2
85 65.8 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 24.278 ug/l
RT: 11.244 min Scan# 1666
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 75 Resp: 72089
Ion Ratio Lower Upper
75 100
77 32.8 19.1 57.5









#77
Bromobenzene
Concen: 19.879 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:156 Resp: 47537
Ion Ratio Lower Upper
156 100
77 157.8 82.8 248.6
158 99.0 49.1 147.4



#78
n-propylbenzene
Concen: 19.502 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 91 Resp: 216566
Ion Ratio Lower Upper
91 100
120 23.3 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.744 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 91 Resp: 132519
Ion Ratio Lower Upper
91 100
126 33.7 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 20.530 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 105 Resp: 163804
Ion Ratio Lower Upper
105 100
120 48.1 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 17.950 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. -0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 75 Resp: 18120
Ion Ratio Lower Upper
75 100
53 111.9 74.5 111.7#
89 47.9 37.9 56.9



#82
4-Chlorotoluene
Concen: 20.186 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 91 Resp: 153784
Ion Ratio Lower Upper
91 100
126 29.1 14.9 44.5









#83
tert-Butylbenzene
Concen: 20.282 ug/l
RT: 11.719 min Scan# 1744
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:119 Resp: 157166
Ion Ratio Lower Upper
119 100
91 61.6 27.9 83.7
134 22.9 11.3 33.9



#84
1,2,4-Trimethylbenzene
Concen: 20.729 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:105 Resp: 164913
Ion Ratio Lower Upper
105 100
120 43.6 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 19.124 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX024128.D
 Acq: 08 Sep 2021 11:11




Instrument :
 MSVOA_X
 ClientSampled :
 VX0908WBS01


Tgt Ion:105 Resp: 187621
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 20.984 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX024128.D
 Acq: 08 Sep 2021 11:11

Tgt Ion:119 Resp: 169438
 Ion Ratio Lower Upper
 119 100
 134 26.8 12.9 38.7
 91 25.4 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 19.474 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. 0.006 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampled :
VX0908WBS01


Tgt Ion:146 Resp: 84255
Ion Ratio Lower Upper
146 100
111 41.3 20.2 60.5
148 63.8 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 19.797 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:146 Resp: 85426
Ion Ratio Lower Upper
146 100
111 40.8 19.6 58.8
148 65.6 31.6 95.0









#89
n-Butylbenzene
Concen: 20.473 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion: 91 Resp: 147603
Ion Ratio Lower Upper
91 100
92 52.0 27.4 82.0
134 26.1 13.2 39.5



#90
Hexachloroethane
Concen: 19.367 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 117 Resp: 25957
Ion Ratio Lower Upper
117 100
201 78.4 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 19.527 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:146 Resp: 81845
Ion Ratio Lower Upper
146 100
111 42.8 20.8 62.2
148 64.7 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.120 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion: 75 Resp: 12444
Ion Ratio Lower Upper
75 100
155 84.0 45.3 135.9
157 103.7 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 20.199 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11




Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01


Tgt Ion:180 Resp: 50955
Ion Ratio Lower Upper
180 100
182 94.6 47.7 143.1
145 30.8 15.0 45.0



#94
Hexachlorobutadiene
Concen: 22.659 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:225 Resp: 25477
Ion Ratio Lower Upper
225 100
223 64.2 31.1 93.2
227 64.6 31.6 95.0









#95
Naphthalene
Concen: 17.976 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Instrument :
MSVOA_X
ClientSampleId :
VX0908WBS01

Tgt Ion:128 Resp: 159514
Ion Ratio Lower Upper
128 100
127 13.1 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 20.075 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX024128.D
Acq: 08 Sep 2021 11:11

Tgt Ion:180 Resp: 50436
Ion Ratio Lower Upper
180 100
182 97.1 47.6 142.9
145 33.3 16.1 48.3

